

HTPD 2018



Contribution ID : 150

Type : not specified

2.5 Microscope requirements to diagnose high-spatial-frequency bright spots in inertial confinement fusion implosions at the national ignition facility

Monday, 16 April 2018 10:45 (120)

Inertial confinement fusion self-emission imaging provides a challenging environment for two-dimensional time resolved x-ray imaging. The short lived (~200 ps) spherical implosion dynamically evolves throughout the deuterium-tritium (DT) compression. Current microscopes with ~10 μm spatial resolution and 20-100 ps time resolution provide sufficient information to infer hot spot volume and emissivity under certain physical constraints. The introduction of high-atomic number materials as shell dopants, in conjunction with the susceptibility of the implosion to seeded hydrodynamic growth, has led to continued observations of high-spatial-frequency x-ray bright spots that evolve internally to the hot DT core. We wish to determine the origin and nature of these features through the application of higher resolution x-ray microscopes. This goal requires addressing both the image forming system and the detector resolution and statistics, in addition to the physics we hope to infer. With new reflective x-ray optics and coded aperture imaging being considered alongside the next generation of fast x-ray detectors, this paper addresses the instrument design requirement to measure 'bright spot' features at the NIF. Prepared by LLNL under Contract DE-AC52-07NA27344. LLNL-ABS-744014.

Primary author(s) : PICKWORTH, Louisa (LLNL)

Presenter(s) : PICKWORTH, Louisa (LLNL)

Session Classification : Session #2, Monday Morning Poster Session