

## HTPD 2018



Contribution ID : 300

Type : not specified

# 14.40 Applications of hybrid-CMOS High-Speed X-ray Framing Cameras on the Z Facility

Thursday, 19 April 2018 10:31 (120)

We have begun using a new generation of high-speed hybrid-CMOS solid-state X-ray framing camera in diagnostics on the Z Facility. Initial applications are in x-ray imaging, spectroscopy, and backlighting. This new camera technology is compact and potentially much simpler to field and operate than traditional microchannel-plate or streak-camera based x-ray imagers. These advantages should make it possible to significantly increase the number of time-resolving diagnostics on Z. A custom electronics system has been developed to operate reliably in the high EMI and radiation environments produced on Z experiments and to survive fielding at a distance as close as 1m from the z-pinch load. We have also developed a compact vacuum interface that enables the sensor to be easily connected to standard vacuum systems for soft x-ray measurement applications. We will describe the performance characteristics of the camera system, the physical dimensions and vacuum interface, computer control and I/O cabling requirements, approaches used to harden the electronics, and initial measurements from experiments on Z. Sandia is a multimission laboratory managed and operated by NTESS LLC, a wholly owned subsidiary of Honeywell Int., Inc., for the U.S. DOE's NNSA under contract DE-NA0003525.

Primary author(s) : STAHOVIK, John (Sandia National Laboratories)

Co-author(s) : COLOMBO, Anthony (Sandia National Laboratories); HARVEY-THOMPSON, Adam (Sandia National Laboratories); JONES, Michael (Sandia National Lab); KELLOGG, Jeff (Sandia National Laboratories); KIMMEL, Mark (Sandia National Laboratories); LOISEL, Guillaume (Sandia National Laboratories); LONG, Joel (Sandia National Laboratories); LOOKER, Quinn (Sandia National Laboratories); SPEAS, Shane (Sandia National Laboratories); WU, Ming (Sandia National Laboratories); CLAUS, Liam (Sandia National Laboratories); ENGLAND, Troy (Sandia National Laboratories); FANG, Lu (Sandia National Laboratories); MITCHELL, Brandon (Sandia National Laboratories); MONTOYA, Andrew (Sandia National Laboratories); ROBERTSON, Gideon (Sandia National Laboratories); SANCHEZ, Marcos (Sandia National Laboratories); ROCHAU, Greg (Sandia National Laboratories); PORTER, John (Sandia National Laboratories); HOHLFELDER, Robert (Sandia National Laboratories); HOHLFELDER, Robert (Sandia National Laboratories); STAHOVIK, John (Sandia National Laboratories)

Presenter(s) : STAHOVIK, John (Sandia National Laboratories); COLOMBO, Anthony (Sandia National Laboratories); HARVEY-THOMPSON, Adam (Sandia National Laboratories); JONES, Michael (Sandia National Lab); KELLOGG, Jeff (Sandia National Laboratories); KIMMEL, Mark (Sandia National Laboratories); LOISEL, Guillaume (Sandia National Laboratories); LONG, Joel (Sandia National Laboratories); LOOKER, Quinn (Sandia National Laboratories); SPEAS, Shane (Sandia National Laboratories); WU, Ming (Sandia National Laboratories); CLAUS, Liam (Sandia National Laboratories); ENGLAND, Troy (Sandia National Laboratories); FANG, Lu (Sandia National Laboratories); MITCHELL, Brandon (Sandia National Laboratories); MONTOYA, Andrew (Sandia National Lab-

oratories); ROBERTSON, Gideon (Sandia National Laboratories); SANCHEZ, Marcos (Sandia National Laboratories); ROCHAU, Greg (Sandia National Laboratories); PORTER, John (Sandia National Laboratories); HOHLFELDER, Robert (Sandia National Laboratories); HOHLFELDER, Robert (Sandia National Laboratories); STAHOVIAK, John (Sandia National Laboratories)

Session Classification : Session #14. Thursday Morning Poster Session