

HTPD 2018



Contribution ID : 304

Type : not specified

14.44 A novel x-ray crystal spectrometer for the diagnosis of high energy density plasmas at the National Ignition Facility

Thursday, 19 April 2018 10:31 (120)

The spectrometer, described in this paper, employs as the Bragg diffracting element a new type of focusing crystal, which is bent in the shape of a torus with a very large major radius, R , and much smaller, minor radii, ρ , which vary in magnitude and their angle of inclination with respect to the major radius, R , along the crystal surface. The source is placed close to the crystal surface to maximize the photon throughput; and the curvature of the crystal is such that, for each wavelength λ , a perfect image of an ideal point source would be obtained in the detector plane, which is at the distance, $R \sin[\Theta(\lambda)]$ from the reflecting crystal points, where $\Theta(\lambda)$ is the Bragg angle for a wavelength λ . The spectral resolution is, therefore, not affected by source-size broadening and can be very high. This spectrometer also provides a high, one-dimensional, spatial resolution perpendicular to the dispersion plane since the crystal-detector distance is much larger than the crystal-source distance [K. W. Hill et al., this conference]. First tests of the spectrometer concept will be conducted with a micro-focus x-ray tube and a Si [220] crystal, which will be glued on 3D-printed substrate. *Supported by U.S. Department of Energy contracts: DE-AC02-09CH11466 and DE-AC52-07NA27344

Primary author(s) : BITTER, M. (Princeton Plasma Physics Laboratory)

Co-author(s) : HILL, K. W. (Princeton Plasma Physics Laboratory); GAO, Lan (Princeton Plasma Physics Laboratory); KRAUS, B. (Princeton Plasma Physics Laboratory); EFTHIMION, P. C. (Princeton Plasma Physics Laboratory); SCHNEIDER, M. (Lawrence Livermore National Laboratory, Livermore); COPPARI, F. (Lawrence Livermore National Laboratory, Livermore); KAUFFMAN, R. (Lawrence Livermore National Laboratory, Livermore); MCPHEE, A. (Lawrence Livermore National Laboratory, Livermore); THORN, D. (Lawrence Livermore National Laboratory, Livermore)

Presenter(s) : BITTER, M. (Princeton Plasma Physics Laboratory); HILL, K. W. (Princeton Plasma Physics Laboratory); GAO, Lan (Princeton Plasma Physics Laboratory); KRAUS, B. (Princeton Plasma Physics Laboratory); EFTHIMION, P. C. (Princeton Plasma Physics Laboratory); SCHNEIDER, M. (Lawrence Livermore National Laboratory, Livermore); COPPARI, F. (Lawrence Livermore National Laboratory, Livermore); KAUFFMAN, R. (Lawrence Livermore National Laboratory, Livermore); MCPHEE, A. (Lawrence Livermore National Laboratory, Livermore); THORN, D. (Lawrence Livermore National Laboratory, Livermore)

Session Classification : Session #14. Thursday Morning Poster Session